

An IMS® Product

## LARGE COLOR CODE INSTRUMENT RINGS



Steam sterilize for at least 4 minutes at 270°F/132°C or 30 minutes at 250°F/121°C. Do not exceed 350°F/177°C. Recommend 30-minute dry time after sterilization cycle.

Manufactured for:  
Hu-Friedy Mfg. Co., LLC  
3232 N. Rockwell St.  
Chicago, IL 60618 | USA  
Hu-Friedy.com

N3074/0318



**50**  
RINGS/BROWN  
IMS-1284L



### INK COLORS



PANTONE  
326



BLACK